Hospital Vaccine Storage Temperature Log (US)

Quality Control | US

WEEK OF	UNIT / LOCATION	DATA LOGGER SERIAL NO.
____ / ____ / _____		

	MON	TUE	WED	THU	FRI	SAT	SUN
Vaccine fridge — AM current °F *36–46°F*							
Vaccine fridge — AM min/max							
Vaccine fridge — PM current °F *36–46°F*							
Vaccine fridge — PM min/max							
Vaccine freezer — AM current °F *?58 to +5°F*							
Vaccine freezer — AM min/max							
Vaccine freezer — PM current °F *?58 to +5°F*							
Vaccine freezer — PM min/max							
DDL reset after reading							
Excursion action taken							
Initials							

Notes

Go digital with Miratag

Skip the paper — fill checklists faster on the Miratag mobile app. Add photos, videos, notes and signatures. Track compliance in real time. Start free at miratag.com

This template is a general-purpose resource, not tailored to any specific jurisdiction. Each organisation must validate compliance with local regulations.

Scan to view